

General Information

Glove Type:	Powder Free Double Polymer Coated, Latex Examination Gloves, Non-Sterile.
Design:	Ambidextrous, Beaded Cuff, Micro-Roughened Finish.
Material:	Natural Rubber Latex.
Colour:	Matt Black.
Protein Content:	Below 50µg/gm glove.
Powder Content:	Below 2mg/glove.
Product Origin:	Made in Malaysia.
Manufactured By:	Barber Healthcare Limited.
Quality Certification(s):	Certified to ISO9001:2000, conforms to EN455 Parts 1, 2 & 3.

Panthera Latex Examination Gloves



Product Dimensions

Glove Size	Product Code	Palm Width (mm)	Minimum Length (mm)		Thickness Single Wall at Palm (mm)	Packing Mode
			EN455 Req's	Our Results		
Petite	PT701	80 ± 5	240	245	0.13 ± 0.03	10x100 gloves/carton
Small	PT702	85 ± 5	240	245	0.13 ± 0.03	10x100 gloves/carton
Medium	PT703	95 ± 5	240	245	0.13 ± 0.03	10x100 gloves/carton
Large	PT704	105 ± 5	240	245	0.13 ± 0.03	10x100 gloves/carton
Extra Large	PT705	115 ± 5	240	245	0.13 ± 0.03	10x90 gloves/carton

Physical Properties

	Before Ageing (minimum)		After Ageing (minimum)	
	EN455 Req's	Our Results	EN455 Req's	Our Results
Force at Break (Newtons)	9.0	10.00	6.0	7.0
Elongation at Break (%)	650	710	500	640

Pre-Shipment Inspection & sampling

Parameter	Characteristics	Inspection Level	AQL	
			EN455 Req's	Our Results
Watertight	Holes	G-1	1.5	0.65
	Dimensions	S-2	4.0	4.0
Physical Properties	Before Ageing/After Accelerated Ageing	S-2	4.0	4.0
	Visual Inspection	Major Defects	G-1	1.5
	Minor Defects	S-4	4.0	4.0

Test & Inspection

Test	Standards
Protein Content	EN455-3 & ASTM D3578 / D5712
Powder Content	ASTM D3578 / D6124
Dimensions/Physical Properties	EN455-2 & ASTM D3578
Watertight	EN455-1 & ASTM D3578 / D5151

Panthera®

PDS-8 : Eng
Revision 2, 09/09

Barber Healthcare Limited, Unit 3, Becksde Court,
Leyburn Business Park, Leyburn, North Yorkshire, DL8 5QA, UK
Telephone: +44 (0) 1969 624 846
Web: www.barberhealthcare.com

Fax: +44 (0) 1969 623 961
eMail: sales@barberhealthcare.com

